

ANALIZA BIAŁEK



Scan for verify



100018011

Laboratory of Analysis

Client: Anonymous
Order number: 100018011
Name: SELLIZA IBUTAMOREN
Number of samples: 1 container

Batch number: 053SRM11
Receive date: 2026-09-03
Start date: 2026-09-11
End date: 2026-09-13

Sample identification

Substance	Method of taking a sample	Form
MK-677	Sample provided by customer	White tablets in a plastic container


Results

Substance	Sample name	Content	Purity
MK-677	SELLIZA IBUTAMOREN 10MG	9,66mg/tab	-----

Research comment

The test sample was applied in the amount of 2 μ l to the BioBasic 8 Thermo Scientific column. During the course of the analysis with the method a linear gradient in the solutions of H₂O and Acetonitrile with the addition of 0.1% trifluoroacetic acid, a peak of the sought substance appeared. Its content, when compared to the reference chromatogram, is 9,66mg/tab. Height the measurement error is +-0,05mg. The actual state of the sample sent before the test is presented on the photo below.

Photo of the test sample




Password

LTBj7JFK

Analiza Bialek sp. z o.o.
KRS: 0000966168
NIP: 8992922970
REGON: 521751875

Headquarters and Laboratory:
ul. Duńska 9, 54-427 Wrocław
Budynek DELTA

Lab technician:
Analysis and research specialist
mgr Arkadiusz Zajac



ANALIZA BIAŁEK




Scan for verify



100018011

Chromatogram



Method

Time	H2O + 0,1% TFA	ACN + 0,1% TFA
0min	95%	5%
2min	95%	5%
12min	5%	95%
15min	5%	95%

Password

LTBj7JFK

Analiza Bialek sp. z o.o.

KRS: 0000966168

NIP: 8992922970

REGON: 521751875

Headquarters and Laboratory:

ul. Duńska 9, 54-427 Wrocław
Budynek DELTA

Lab technician:

Analysis and research specialist
mgr Arkadiusz Zajac